

0 "zero" is on the top

Read volumes directly from the scale, which is from top to bottom. Notice that 5.7 mL is **below** 5.0 mL and **not above** 5.0.

Make sure there is enough reagent in the buret before beginning each titration.

50 mL is on the bottom

Close the stopcock when filling the first time [horizontal position]

Eliminate bubbles from tip before starting the first titration